

Vibrating Sample Magnetometer (VSM)

Building 26 (PEMS South), room G40, contact Wayne Hutchison, Annemieke Mulders

PAR Model 155 magnetometer coupled with:

- closed cycle refrigerator (8 – 300 K)
- electromagnet (0 – 1 T)
- LabVIEW data access

